

DESCRIPTION:

Hall Probe G for F1A Magnetic Transducers with the naked Si chip is a very thin (0.5mm) and long (42mm) single-chip fully integrated 1-axis Hall-Probe. It measures magnetic fields perpendicular to the probe plane (By).

The Hall Probe G for F1A Magnetic Transducers contains a CMOS integrated circuit, which incorporates a group of four Hall elements, biasing circuits, amplifiers, and a temperature sensor. The integrated Hall elements occupy very small area ($0.4 \times 0.04\text{mm}^2$), which provides very high spatial resolution of the probe.


The on-chip application of the spinning-current technique in the biasing of the Hall elements suppresses the planar Hall effect. The signal pre-processing on the chip enables a very high frequency bandwidth (DC to 75 kHz) of the probe; and on-chip signal amplification provides high output signals of the Hall probe.

The sensor chip is embedded in the probe package and connected to the CaH cable.



The output of the Hall Probe are high-level analog voltages proportional with each of the measured components of a magnetic flux density and a voltage proportional with the probe temperature.

KEY FEATURES:

- **The G Hall probe with the naked Si chip is very thin (0.5mm) and long (42mm)**
- **Fully integrated CMOS 1-axis (By) Hall Probe**
- **Very high spatial resolution ($0.40 \times 0.01 \times 0.04\text{mm}^3$)**
- **High frequency bandwidth (from DC up to 75kHz)**
- **Virtually no planar Hall Effect**
- **Negligible inductive loops on the Probe**
- **Integrated temperature sensor on the probe for temperature compensation**
- **The Hall Probe G for F1A Magnetic Transducers is consisting part (Module H) of the F1A Magnetic Field Transducer, Digital Teslameter and Magnetic Field Mapper**
- **Suitable as Y-axis (0Y) Hall Probe**
- **Available as a separate unit for OEM customers**



PROBE DIMENSIONS AND CHARACTERISTICS:



Probe and Cable Dimensions & Tolerances							
Dimension	[mm]	Dimension	[mm]	Dimension	[mm]	Dimension	[mm]
A	2.0 ± 0.1	G	4.8 ± 0.2	M	50 ± 1	S	2.8 ± 0.1
B	4.0 ± 0.1	H	16 ± 1	N	35 ± 3	T	0.2 ± 0.1
C	2.0 ± 0.1	I	0.5 ± 0.1	O	30 ± 1	U	0.15
D	1.5 ± 0.2	J	2.0 ± 0.1	P	50 ± 2	V	0.64
E	3.8 ± 0.2	K	42.3 ± 0.2	Q	XX ± 1% ¹⁾	W	1.1 ± 0.1
F	6.0 ± 0.2	L	10.0 ± 0.1	R	6.0 ± 0.1		

Figure 1: Dimensions of G Hall probe and Cable (Module H)




Figure 2: The reference Cartesian coordinate system of G Hall probe

Dimension	X [mm]	Y [mm]	Z [mm]
Magnetic field sensitive volume (MFSV)	0.40	0.01	0.04
Position of the center of MFSV (corresponding to MFSP, see Fig.1)	1.0 ± 0.1	-0.4 ± 0.05	-0.35 ± 0.1
	<ul style="list-style-type: none"> • 2.0 ± 0.1 - Longer, thinner part 	<ul style="list-style-type: none"> • 0.5 ± 0.1 - Longer, thinner part 	<ul style="list-style-type: none"> • 42.3 ± 0.2 - Longer, thinner part
Total Probe external dimensions	<ul style="list-style-type: none"> • 4.0 ± 0.1 Shorter, thicker part of the probe 	<ul style="list-style-type: none"> • 2.0 ± 0.1 - Shorter, thicker part of the probe 	<ul style="list-style-type: none"> • 12.0 ± 0.1 - Shorter, thicker part of the probe
Angular accuracy of the axes	± 0.5° with respect to the reference surface		
CaH Cable	Shielded, with a flexible thin part near the probe		
Total length of the CaH cable:	<ul style="list-style-type: none"> • Standard: 2 m (Probe notation: OYG02) • Optional: xx m (Probe notation: OYGxx) <p><i>Note: Different lengths are available upon request.</i></p>		

INSTALLATION MANUAL FOR G HALL PROBE

Warning: the Probe Tip is Fragile!

The probe package is made of ceramics, and can be easily broken. Therefore, avoid any mechanical contact of the probe tip with other objects! Moreover, avoid the immersion of the probe of any liquid, and its exposure to moisture and aggressive gasses.

The following precautions shall help ensure that the transducer accurate calibration remains preserved:

The mounting of the probe should be carried out by application of very low pressure to its back-end and thin wires. If the probe head is clamped, the user should make sure that the substrate surface in contact with the reference plane of the probe is flat and covers as much of the probe reference surface as possible. Do not apply more force than required to hold the probe in its place.

In order to prevent rupture of the thin wires from the probe head, the user should fix and secure the probe cable in the proximity of the probe. The thin wires of the flexible section of the cable may be folded with care; repeated strong bending should be avoided.