



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

GMW ASSOCIATES
 955 Industrial Road
 San Carlos, CA 94070
 Sandro Renteria Phone: 650 802 8292

CALIBRATION

Valid To: March 31, 2026

Certificate Number: 4349.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,6}:

I. Electrical – DC/Low Frequency

Parameter/Range ⁵	Frequency	CMC ² (±)	Comments
DC Ratio Error ³ – Current & Voltage Output (15 to 600) A 600 A to 11 kA		0.2 % of ratio 0.1 % of ratio	Calibration of DC current transducers using reference transducer comparison method
AC Ratio Error ^{3,4} – Current & Voltage Output 5A to 2kA (2 to 8) kA 5A to 8kA	(5 to 100) Hz (5 to 100) Hz (100 to 400) Hz	0.05 % of ratio 0.1 % of ratio 0.2 % of ratio	Calibration of AC current transducers using reference transducer comparison method
AC Phase Displacement ³ – Current & Voltage Output 5A – 8kA	(5 to 100) Hz (100 to 400) Hz	0.09 ° 0.11 °	Calibration of AC current transducers using reference transducer comparison method

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ AC Current Range values are RMS (sinusoidal) values. Limited to 2 kA at 400 Hz.

⁵ The current range refers to the Transducer's Nominal Current.

⁶ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

GMW ASSOCIATES

San Carlos, CA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 20th day of February 2024.

A blue ink signature of Mr. Trace McInturff, written in a cursive style.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4349.01
Valid to March 31, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.